

## X. Workshop on Resistive Plate Chambers and Related Detectors

Contribution ID: 73

Type: **not specified**

# VME based DAQ system for INO RPC stack

*Wednesday, 10 February 2010 17:00 (2 hours)*

A 12 layer 1m x1m glass RPC stack is operational and collecting cosmic ray data for the last two years at TIFR. A VME based DAQ system with various graphical interfaces for data acquisition and monitoring has been developed. Both hardware and software aspects of this system will be discussed in this paper.

**Primary author:** Mr DEEPAK, Samuel (Tata Institute of Fundamental Research, Mumbai)

**Presenter:** Mr DEEPAK, Samuel (Tata Institute of Fundamental Research, Mumbai)

**Session Classification:** Poster session