

ESR Status August 28, 2018



- vacuum pressure good for commissioning (average 5×10^{-11} mbar)
- a leak in the injection line was repaired and section was baked out again
- all mechanical work in the ESR will be finished today
- tomorrow preparation of devices for dry run
- analog signal preparation for beam commissioning
- mechanical work and cabling of electron cooler is finished, cathode preparation and HV conditioning started this week
- preparation of control system is continued