

MM 24.09.2019 - PSU

- **Alvarez_2.0:**
 - MSP plan put on server
 - started process of refining human resources for (de)installation & commissioning
- **FoS:**
 - SAT (repetition of FAT after transp.)
 - all ok, except ...
 - vacuum: Cu seals -> to be checked
 - low level rf-measurements started
 - empty cavity: first modes measured as predicted
 - first drift tube & stem transferred to GALVANO-T for Cu-plating
 - today: kick-off at DANFYSIK for tube with quad
- **Safety**
 - List of “prüfungspflichtige Arbeitsmittel” sent to M. Hirsch incl. potential issues -> awaiting response

